

Order No.
81 01321


Water Vapour 0.1/a

Standard Measuring Range : 0.1 to 1.0 mg/L
Number of Strokes (n) : 3
Time for Measurement : app. 1.5 min
Standard Deviation : ± 15 to 20 %
Colour Change : yellow \rightarrow blue

Ambient Operating Conditions

Temperature : 0 to 30 °C

Reaction Principle



Cross Sensitivity

Generally basic substances cause plus errors and acidic substances cause minus errors.

No interference by:

- 1,200 ppm nitrogen dioxide
- 6,000 ppm sulphur dioxide
- 3,000 ppm hydrogen sulphide
- 2,000 ppm ethanol
- 2,000 ppm acetone

Additional Information

The first scale mark corresponds to 0.05 mg H₂O / L.



SI-184-2001