


Tube No. **NX145**



Nexteq
Short Term



1. PERFORMANCE


- 1) Measuring range : 2-30 ppm 1-5 ppm
- Number of pump strokes : 1 (100mL) 2 (200mL)
- 2) Sampling time : 1 minute / 1 pump stroke
- 3) Detectable limit : 0.2 ppm (200mL)
- 4) Shelf life : 2 years
- 5) Operating temperature : 10~40°C
- 6) Temperature compensation : Necessary (0~20°C) (See "TEMPERATURE CORRECTION TABLE")
- 7) Reading : Direct reading from the scale calibrated by 1 pump stroke
- 8) Color change : Pale purple→Pale yellow

2. RELATIVE STANDARD DEVIATION

RSD-low: 10% RSD-mid.: 5% RSD-high: 5%

3. CHEMICAL REACTION

By reacting with alkali, Amine is produced. Further, pH indicator is discolored by reacting together with Phosphoric Acid.



4. CALIBRATION OF THE TUBE

DIFFUSION TUBE METHOD

5. INTERFERENCE AND CROSS SENSITIVITY



Substance	Interference	ppm	Coexistence
Sulfur Dioxide FIG.1	The accuracy of readings is not affected.	200	Lower readings are given.
Carbon Dioxide FIG.2	"	0.1%	"
Chlorine	"	—	"
Ammonia	Similar stain is produced.	—	Higher readings are given.
Amines	"	—	"
Hydrazine	"	—	"

(NOTE)

When the concentration is below 5 ppm, 2 pump strokes can be used to determine the lower concentration.

Following formula is available for the actual concentration.

Actual concentration = 1/2 × Temperature corrected value



TEMPERATURE CORRECTION TABLE

Tube Readings (ppm)	Corrected Concentration (ppm)		
	10 °C (50 °F)	15 °C (59 °F)	20-40 °C (68-104 °F)
30	81	40	30
25	67	33	25
20	54	27	20
15	40	20	15
10	27	13	10
5	13	7	5
2	5	3	2