



Standard Measuring Range : 20 to 100 ppm
Number of Strokes (n) : 10
Time for Measurement : app. 2.5 min
Standard Deviation : ± 10 to 15 %
Colour Change : white \rightarrow yellow brown

Ambient Operating Conditions

Temperature : 0 to 50 °C
Absolute Humidity : 3 to 30 mg H₂O / L

Reaction Principle

- a) $2 \text{R-SH} + \text{Cu}^{2+} \rightarrow \text{Cu}(\text{RS})_2 + 2 \text{H}^+$
b) $\text{Cu}(\text{RS})_2 + \text{S} \rightarrow$ yellow brown copper compound

Cross Sensitivity

Higher molecular weight alkyl mercaptans (e.g. propyl- and butyl-mercaptans) are indicated with approximately the same sensitivity.

Hydrogen sulphide is indicated with approximately twice the sensitivity of the mercaptans (e.g. 10 ppm hydrogen sulphide gives an indication of 20 ppm).

Additional Information

After performing the required ten pump strokes the reagent ampoule must be broken. The liquid of the ampoule must be transferred to the indicating layer and carefully drawn through it using the pump.

After completing measurement wait for 3 min prior to evaluation.