


Tube No. **NX156M**



1. PERFORMANCE


- | | | |
|--------------------------|---|------------|
| 1) Measuring range | : 5-80 ppm | 2.5-40 ppm |
| Number of pump strokes | : 1/2 (50mL) | 1 (100mL) |
| 2) Sampling time | : 30 seconds / 1/2 pump stroke | |
| 3) Detectable limit | : 1 ppm (100mL) | |
| 4) Shelf life | : 2 years | |
| 5) Operating temperature | : 0~40°C | |
| 6) Reading | : Direct reading from the scale calibrated by 1/2 pump stroke | |
| 7) Color change | : Yellow→Pink | |

2. RELATIVE STANDARD DEVIATION

RSD-low: 15% RSD-mid.: 10% RSD-high: 5%

3. CHEMICAL REACTION

By reacting with Mercuric Chloride, Hydrogen Chloride is produced and pH indicator is discolored.



4. CALIBRATION OF THE TUBE

STANDARD GAS CYLINDER METHOD

5. INTERFERENCE AND CROSS SENSITIVITY

Substance	Interference	Coexistence
Hydrogen Sulfide	Similar stain is produced.	Higher readings are given.
Phosphine	"	"
Other Mercaptans	"	"
Arsine	"	"
Hydrogen Selenide	"	"
Hydrogen Cyanide	"	"
Nitrogen Dioxide	The accuracy of readings is not affected.	Lower readings are given.
Ammonia	"	"
Sulfur Dioxide	"	"

(NOTE)

In case of 1 pump stroke, following formula is available for the actual concentration.

Actual concentration = 0.5 × Reading value